

7-AUGUST-2015

URGENT: Field Safety Notice

FSCA identifier: Product Field Action RA2014-096
Type of Action: Field Safety Corrective Action: Correction
Description: RIO Base Array
Catalog #: 112220
Lot Code: All

Dear Distributor/ Risk Management:

Incidences have been reported where the locking pin on the RIO Base Array (P/N 112220) may be damaged during assembly onto the RIO Base Array Connector (P/N 205143). The incidence of this issue is uncommon (only 0.15% of cases over the past 36 months) and, of the reported cases, all were successfully completed with no reported injury to the patient or OR staff.

While the incidence of this issue is uncommon, it does potentially affect all sites with a RIO system. We are therefore communicating this issue to all affected customer sites through letters to the OR Administrators. Our engineering team has provided a solution to this issue and a new RIO Base Array Connector and Base Array will be installed on the RIO system during regularly scheduled product maintenance.

The RIO system can continue to be used while this action is in process, as the issue is readily identified during RIO set up.

Our records indicate that you have received the above referenced product. Please assist us in meeting our regulatory obligation by:

1. Please complete the attached acknowledgement form indicating that you have received this Important Medical Device Correction letter, and return it to us.
2. The RIO system can continue to be used. The incidence of locking pin damage is uncommon (0.15% of cases) and, in the event of occurrence, readily detectable, and can be timely remediated by replacing the [locking pin] with one taken from another RIO Base Array. Ensure that a second RIO Base Array is available during the pre-surgery check to remediate and provide a timely solution should the uncommon issue occur and, in the event of occurrence, return the damaged unit to MAKO Surgical Corp.
3. Complete the attached customer response form and return the form and any affected devices to your local Stryker Representative.

(Please complete this form even if you do not have any product to return. This will preclude the need to Stryker to send any reminder notice)

6. Please inform Stryker of any adverse events associated with the use of the subject devices.
 - a. Comply with any local regulations concerning the reporting of adverse events to local Competent Authorities.


Stryker® Orthopaedics maintains its commitment to developing, manufacturing and marketing the highest quality products for surgeons and patients. We apologize for any inconvenience this Field Safety Corrective Action may create and appreciate your cooperation with our request.

If you have any further enquiries, please do not hesitate to contact the undersigned.

Yours Sincerely,

STRYKER® ORTHOPAEDICS
FIELD SAFETY CORRECTIVE ACTION ACKNOWLEDGMENT FORM

August 7, 2015



FSCA identifier: Product Field Action - RA2014-096

Description RIO Base Array

Catalog #: 112220

Lot Code: All

Type of Action: Correction

I have received the notification from Stryker® Orthopaedics dated August 7, 2015 stating that they initiated a Field Safety Corrective Action of the above referenced product.

Representative
(Signature)

Date

Representative
(Print)

Please fax this signed and dated form to **1-855-251-3635**