

**URGENT  
PRODUCT RECALL**

January 17, 2018

Product Field Action Number: 1707484

Description: Triathlon TS Baseplates Sizes 3 and 4

Affected Catalog Number(s): 5521-B-300, 5521-B-400

Affected Lot Number(s): ATY4OA, ATV7IA

Dear Customer,

Stryker Orthopaedics (“Stryker”) has initiated a voluntary, lot-specific recall for the Stryker knee components referenced above. The intent of this letter is to list known hazards potentially associated with the use of the implant and list any risk mitigation factors.

Issue:

Stryker has discovered that the two product/lot combinations referenced above may contain the incorrect size implant from what is labeled on the box.

Potential Hazards:

Technical and medical assessments are currently underway to determine any potential hazards associated with the use of the product. Additional communication will be forwarded upon completion of the internal investigation on this issue.

Risk Mitigation:

The Triathlon TS baseplates are laser marked with lot and catalog number. Although the product packaging for these two lots may not match the product contained within, the laser marked details on the implant would increase the likelihood that the surgeon or surgical staff would recognize that the incorrect implant was contained in the package.

Actions Needed

1. Please inform users of this Urgent Product Recall and forward this notice to all those individuals who need to be aware within your organization.
2. Hospitals/Branches/Agencies: Complete and sign the Recall Notification Business Reply Form which you will receive via UPS and fax a copy to 1-855-251-3635 or email to [SO M Product Field Action Response@stryker.com](mailto:SO M Product Field Action Response@stryker.com)
3. Hospitals/Branches/Agencies: Return all affected products from your inventory locations to the following address.

**Stryker Orthopaedics/PFA Product Returns  
Attn: Distribution Inventory Team  
325 Corporate Drive  
Dock M-East  
Mahwah, NJ 07431  
Ref. PFA 1570495**

Our records indicate that you have received the above referenced implant. It is our responsibility to ensure that customers who may have received this affected implant also receive this important communication.

**Please assist us in meeting our regulatory obligation by faxing back the attached Recall Notification Business Reply Form within 5 days.**

We regret any inconvenience this action may cause you and if you have any questions, feel free to contact me at (201) 831-6693.

Sincerely,

Eric Petschler  
Manager, Divisional Regulatory Compliance

**STRYKER ORTHOPAEDICS  
URGENT MEDICAL DEVICE RECALL  
NOTIFICATION BUSINESS REPLY FORM**

January 17, 2018

Product Field Action Number: 1707484

Description: Triathlon TS Baseplates Sizes 3 and 4

Affected Catalog Number(s): 5521-B-300, 5521-B-400

Affected Lot Number(s): ATY40A, ATV71A

I have received the product recall letter from Stryker Orthopaedics dated January 17, 2018 stating that the company has initiated a voluntary, lot-specific recall of the above referenced implant.

<b>We have not located any of these devices in our inventory:</b> <i>(please delete if not applicable)</i>				
<b>We have located the following devices</b> <i>(please include additional lines as necessary):</i>				
Product	Product Reference	Lot Number(s)	Qty	Qty Quarantined
5521-B-300	TRI TS BASEPLATE SIZE 3	ATV71A		
5521-B-400	TRI TS BASEPLATE SIZE 4	ATY40A		

\_\_\_\_\_  
Hospital or Stryker Branch Name

\_\_\_\_\_  
Date

\_\_\_\_\_  
Hospital/Agent/Risk Rep or Stryker Branch Rep  
(Signature)

**PLEASE COMPLETE THIS FORM WITHIN 5 BUSINESS DAYS AND RETURN IT BY USING THE EMAIL OR FAX LISTED BELOW:**

Email: [SO M Product Field Action Response@stryker.com](mailto:SO M Product Field Action Response@stryker.com).

Fax: 1-855-251-3635