

B. Braun Melsungen AG
Division Aesculap
Vascular Systems

Sieversufer 8
12359 Berlin

Clinic

Contact person: Dr. Bernhard Jänicke

Administration
Street
City
Country

Telephone: +49 30 568207-170
Fax: +49 30 568207-123
Email: bernhard.jaenicke@bbraun.com
Internet: <http://www.bbraun.com>

Date: May 29, 2015

Dear customer,

We are aware that you are one of the customers who use in your facility Coroflex ISAR size **2.5 x 14 mm** with the article no. **5028712** LOT **15A20809** in the interventional cardiology.

Due to a recent complaint we reviewed our production process and recognized that we have to assume that one product of the named LOT may be faulty regarding its packaging. In one individual case the label shows a different stent diameter and length compared to the packed stent. This product LOT must not be used anymore.

B. Braun representatives will contact you in a timely manner to retrieve the goods of the mentioned article number on stock. The goods shall be returned to

B. Braun Melsungen AG Vascular Systems
Sieversufer 8
12359 Berlin

by **15th of June 2015**.

Our B. Braun representative will offer you appropriate products as replacements according to the factual situation in order to guarantee the maintenance of product provision.

We apologize for any inconveniences this process of safekeeping may cause and appreciate very much your understanding and support.

B. Braun Melsungen AG

Senior Vice President
Vascular Systems

Regulatory Affairs Manager

Vors. des Aufsichtsrats:
Prof. Dr. h.c. Ludwig Georg Braun

Vorstand:
Prof. Dr. Heinz-Walter Große
(Vorsitzender)
Dr. Annette Beller
Otto Philipp Braun

Prof. Dr. Hanns-Peter Knaebel
Dr. Meinrad Lugan
Caroll H. Neubauer, LL.M.
Markus Strotmann

Sitz der Gesellschaft: Melsungen
Reg. Gericht: Amtsgericht Fritzlar
HRB 11 000