



URGENT MEDICAL DEVICE CORRECTION

GE Healthcare

Healthcare Systems
9900 Innovation Drive
Wauwatosa, WI 53226
USA

<Date of Letter Deployment>

GEHC Ref# 60878

To: Hospital Administrators / Risk Managers
Radiology Department Managers
Radiologists

RE: Software Version: GE Healthcare MRI systems.

GE Healthcare has become aware of a potential safety issue due to software versions not being maintained properly at some sites.

Safety Issue It has come to our attention that the latest on-site software version may not have been reinstalled at some sites after service activities were performed that required software to be reloaded.

The most recently deployed on-site software version may include safety updates.

Safety Instructions None. No User action required. GE Healthcare will inspect all affected systems and ensure the proper software version is installed.

Affected Product Details This correction applies to some units of the following GE Healthcare MRI system types: Discovery MR450, Discovery MR750, Signa HDx, Signa HDxt, Signa HDi, GE 1.5T Signa HDe, Optima MR360, Brivo MR355, Signa Excite-HD 1.5T & 3T, Signa Excite 1.5T HD Twinspeed, Signa Excite 1.5T HD EchoSpeed, Signa Excite 1.5T HD Highspeed, 1.5T Signa Infinity TwinSpeed , 1.5T Signa Infinity EchoSpeed Plus, 1.5T Signa Infinity HiSpeed Plus, 1.0T Signa Infinity HiSpeed Plus. 1.0T Signa Infinity SmartSpeed, Signa EXCITE 3.0T, Signa EXCITE 3.0T HD, Signa Excite 1.5T TwinSpeed, Signa Excite 1.5T EchoSpeed, Signa Excite 1.5T HiSpeed, Signa Excite 1.5T SmartSpeed, Signa Excite 1.5T, Signa Excite 3.0T, Signa Contour/I, Signa OpenSpeed, Vectra, MR Max, Signa Profile, Signa Ovation, Optima MR450w GEM, Discovery MR750W GEM, Optima MR450w, Signa PET/MR, MR950.


Product Correction GE Healthcare will inspect all affected systems and ensure the proper software version is installed. This will be performed at no cost to you. GE Healthcare will check systems remotely (on-line) where possible, and will visit your site if a remote check is not possible.

GE Healthcare will advise you if a correction is needed, or if a visit to your site is needed. If a correction is needed, a service representative will contact you to arrange for this correction.

Contact Information If you have any questions concerning this notification, contact your local service representative or call 800-437-1171 (available 24 hours a day).

Please be assured that maintaining a high level of safety and quality is our highest priority. If you have any questions, please contact us immediately per the contact information above.

Sincerely,



James Dennison
Vice President QARA
GE Healthcare Systems

Werner Van den Eynde, M.D.
Acting Chief Medical Officer
GE Healthcare